



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 Chemistry Reference Sensor
Lot Number 26041053

Part Number 57466
Exp Date 2027-08-10
Date of Manufacture 2026-02-10

| | <u>Spec*</u> | <u>Pass/Fail</u> |
|----------------|---------------|------------------|
| Na Slope | 8.70 - 11.80 | Pass |
| BioPro Ctl L 1 | 70.6 - 86.2 | Pass |
| BioPro Ctl L 3 | 164.7 - 201.3 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

* Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

3/4/2026
Date