



**CERTIFICATE OF ANALYSIS**

STP pHox Ultra/CCX Lactate Membrane  
Lot Number 26050021

Part Number 48554  
Exp Date 2027-05-19  
Date of Manufacture 2026-02-19  
UDI# (01)00385480485548(11)260219(17)270519(10)26050021

	<u>Spec*</u>	<u>Mean Result</u>
Slope	≥ 15.00	37.05
pHox Ultra/CCX Ctl L4	1.7 - 2.3	2.0
pHox Ultra/CCX Ctl L5	6.3 - 7.7	7.3
Lin Soln Set A Level 1	14.0 - 20.0	18.4

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

3/10/2026  
Date