



## CERTIFICATE OF ANALYSIS

Stat Profile pHox Ultra / CCX BUN Membrane Kit  
Lot Number 26036020

Part Number 48596  
Exp Date 2027-05-05

Date of manufacture 2026-02-05

UDI# (01)00385480485968(11)260205(17)270505(10)26036020

	<u>Spec</u>	<u>Mean Result</u>
BUN Slope	9.30 - 12.90	12.25
CCX L 4	14 - 24	19
CCX L 5	49 - 63	56

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink that reads "Olivia Rammer".

Approval

QC Inspector

Title

3/3/2026

Date