



CERTIFICATE OF ANALYSIS

Bioprofile Control Level 1
Lot Number 600913

Catalog Number 22600
Exp Date 2027-07

		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.600 - 6.800	Pass
PO2	<i>mmHg</i>	53 - 63	Pass
PCO2	<i>mmHg</i>	57 - 67	Pass
Gluc	<i>g/L</i>	0.85 - 1.15	Pass
Lac	<i>g/L</i>	0.85 - 1.15	Pass
NH4+	<i>mmol/L</i>	0.65 - 1.15	Pass
Na+	<i>mmol/L</i>	72.0 - 82.0	Pass
K+	<i>mmol/L</i>	1.85 - 2.15	Pass
iCa++	<i>mmol/L</i>	0.55 - 0.75	Pass
PO4	<i>mmol/L</i>	3.9 - 4.4	Pass
Osmo	<i>mOsm/Kg</i>	213 - 219	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH, PO2, PCO2, Lac, Gluc, NH4, Na, K, iCa.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

03/03/2026
Date