



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA	
Description:	TEST STRIPS, STATSTRIP HEMOGLOBIN/HEMATOCRIT, PACKAGED 50 CT	
Catalog Number:	58786	
Nova Biomedical P/N	58786	
<u>Target Range (Hct)</u>	<u>CV (%)</u>	
28-36 %	≤ 8%	PASS / FAIL
38-46 %	≤ 8%	PASS / FAIL
48-56%	≤ 8%	PASS / FAIL
Qty:	10,008	
Lot No.	1526057249	
Date of Mfg. (DOM):	2026/02/26 (YYYY/MM/DD)	
Expiration Date:	2028/02/26 (YYYY/MM/DD)	
Unique Device Identifier (UDI):	(01)20385480587864 (11)260226 (17)280226 (10)1526057249	

Maureen Broome 3/11/26
 Quality Control Manager Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454

Released Confidential