



CERTIFICATE OF ANALYSIS

STP pHox Ultra / CCX Chloride Sensor
Lot Number 26009021

Part Number 48557
Exp Date 2027-04-09
Date of Manufacture 2026-01-09
UDI# (01)10385480485576(11)260109(17)270409(10)26009021

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	8.50 - 13.00	Pass
pHOx Ctl L4	121.3 - 130.3	Pass
pHOx Ctl L5	91.5 - 100.5	Pass
Std C mV	10.00 - 90.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

2/9/2026
Date