



**CERTIFICATE OF ANALYSIS**

STP Prime ABG Cal Cartridge 500 Sample  
Lot Number 26049067

Ref Number 53464  
Exp Date 2027-07-21  
Date of Manufacture 2026-01-21  
UDI# (01)00385480534642(11)260121(17)270721(10)26049067

Cal A	L/N	26036046	
		<u>Assigned Value*</u>	<u>Result</u>
	pH	7.356	N/A
	pCO2 mmHg	25.1	N/A

Cal B	L/N	26036047	
		<u>Spec</u>	<u>Pass/Fail</u>
	pH	6.837 - 6.843	Pass
	pO2 mmHg	167.0	N/A

Flush	L/N	26021032	
		<u>Assigned Value*</u>	<u>Result</u>
	pCO2 mmHg	52.4	N/A
	pO2 mmHg	103.8	N/A

Reference	L/N	26023013	
		<u>Spec</u>	<u>Pass/Fail</u>
	KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Auishi.M*

Approval

QC Inspector  
Title

03/12/2026  
Date