



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA		
Description:	TEST STRIPS, STATSTRIP LACTATE, PACKAGED		
Catalog Number:	47486		
Nova Biomedical P/N	47486		
<u>Target Range</u>	<u>CV (%)</u>	<u>SD (mmol/L)</u>	
0.8-2.5 mmol/L	≤ 6.5%	±0.3	<input checked="" type="radio"/> PASS <input type="radio"/> FAIL
3.0-5.0 mmol/L	≤ 5.5%	±0.5	<input checked="" type="radio"/> PASS <input type="radio"/> FAIL
10-14 mmol/L	≤ 5.0%	±1.2	<input checked="" type="radio"/> PASS <input type="radio"/> FAIL
Qty:	846		
Lot No.	2626061249		
Mfg. Date:	2026-03-02		
Expiration Date:	2028-03-02		
Unique Device Identifier (UDI):	(01)20385480 <u>474861</u> (11) <u>260302</u> (17) <u>280302</u> (10) <u>2626061249</u>		

MacBryne
Quality Control Manager

3/20/26
Date

Nova Biomedical Corporation . 200 Prospect Street Waltham, MA 02454