



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, STATSTRIP A CREATININE, 50 COUNT
Catalog Number:	58788
Nova Biomedical P/N	58788
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	PASS/FAIL PASS/FAIL PASS/FAIL
Qty:	100
Lot No.	1325314129
Mfg. Date:	2025-11-10
Expiration Date:	2026-11-10
Unique Device Identifier (UDI):	(01)20385480587888 (11) <u>25 1110</u> (17) <u>26 1110</u> (10) <u>1325314129</u>

Quality Control Manager

3/23/26
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454