



CERTIFICATE OF ANALYSIS

Stat Profile pHox/Basic Calibrator Cartridge B
Lot Number 26041063

Ref Number 23937
Exp Date 2027-07-29

Date of Manufacture 2026-01-29

UDI# (01)00385480239370(11)260129(17)270729(10)26041063

Cal A L/N 26036045

Cal B L/N 26035010

Cal C L/N 26035011

HCO3 mmol/L 20.1 - 20.9 Pass
Na mmol/L 139.3 - 141.3 Pass

HCO3 mmol/L 40.2 - 41.8 Pass

pH @ 37C 7.342 - 7.348 Pass

Cal D L/N 26035012

Reference L/N 26029015

pH @ 37C 6.792 - 6.798 Pass
Na mmol/L 71.5 - 72.5 Pass

KCl mol/L 1.8 - 2.2 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi-M

Approval

QC Inspector
Title

03/2/2026
Date