



CERTIFICATE OF ANALYSIS

Stat Profile pHox Ultra / CCX SO2 Sensor
Lot Number 26041008

Part Number 48591
Exp Date 2029-02-10
Date of manufacture 2026-02-10
UDI#(01)00385480485913(11)260210(17)290210(10)26041008

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	7.00 - 14.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Sean Quinn".

Approval

QC Inspector
Title

2/26/2026
Date