



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	STATSENSOR-I CREAT Test Strips
Catalog Number:	P/N 43272
Nova Biomedical P/N	P/N 43272
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	PASS/ FAIL PASS/ FAIL PASS/ FAIL
Qty:	2,718
Lot No.	4926063129
Mfg. Date:	2026-03-04
Expiration Date:	2027-03-04
Unique Device Identifier (UDI):	(01)20385480432720 (11)260304 (17)270304 (10)4926063129

Maureen
 Quality Control Manager

3/18/26
 Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454