



## CERTIFICATE OF ANALYSIS

BioPro Flex2 ESM Reagent Cartridge  
Lot Number 26072020

Ref Number 56227  
Exp Date 2027-02-18

Solution S L/N 26050027

	<u>Spec</u>	<u>Pass/Fail</u>
pH	6.900 - 7.100	Pass

Solution DP L/N 26049064

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl %	0.43 - 0.73	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "J. Smith", is located above the approval text.

Approval

QC Inspector  
Title

3/31/2026  
Date