



**CERTIFICATE OF ANALYSIS**

Reagent Pack Autosampler Flex 2  
Lot Number 26044071

Part Number 62450  
Exp Date 2027-01-22  
Date of Manufacturing 2026-01-22

Solution DP L/N		26022025	Solution SF L/N		26028100
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
NaOCI %	0.43 - 0.73	Pass	pH	6.900 - 7.100	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, HCO3, pH, KCl, iCa, NH4, K.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

3/20/2026  
Date