



**CERTIFICATE OF ANALYSIS**

STP pHoxUltra/CCX CO-Oximeter Cal Cartridge  
with Bilirubin and Depro Solution  
Lot Number 26009015

Ref Number 48559  
Exp Date 2027-01-09  
Date of Manufacture 2026-01-09  
UDI# (01)00385480485593(11)260109(17)270109(10)26009015

Dil L/N	26009015		Flush L/N	26009015	
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>
Surfactant	N/A	Present	Surfactant	N/A	Present
Buffers	N/A	Present	Preservatives	N/A	Present
Preservatives	N/A	Present			

Depro L/N	26009015		tHb Calibrator L/N	26080001	
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>
NaOCl %	2.90 - 3.20%	Pass	g/dL	16.0	
			g/L	160	
			mmol/L	9.9	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. Total Hemoglobin (THb) is traced by using Cyanmethemoglobin method.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

4/1/2026  
Date