



CERTIFICATE OF ANALYSIS

STAT PROFILE pHox Ultra/CCX Potassium Sensor
Lot Number 26035065

Ref Number 48582

Exp Date 2027-05-04

Date of Manufacture 2026-02-04

UDI# (01)00385480485821(11)260204(17)270504(10)26035065

| | <u>Spec</u> | <u>Pass/Fail</u> |
|----------|---------------|------------------|
| Slope | 9.00 to 11.00 | Pass |
| Std C mV | 0 ± 50.00 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

3/17/2026
Date