



**CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	STATSENSOR-I CREAT Test Strips
Catalog Number:	P/N 43272
Nova Biomedical P/N	P/N 43272
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	<del>PASS</del> / FAIL <del>PASS</del> / FAIL <del>PASS</del> / FAIL
Qty:	2,250
Lot No.	4926065129
Mfg. Date:	2026-03-06
Expiration Date:	2027-03-06
Unique Device Identifier (UDI):	(01)20385480432720 (11) <u>260306</u> (17) <u>270306</u> (10) <u>4926065129</u>

Max Bowne  
Quality Control Manager

3/23/26  
Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454