



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 Chemistry Reference Sensor
Lot Number 26071036

Part Number 57466
Exp Date 2027-09-12
Date of Manufacture 2026-03-12

	<u>Spec*</u>	<u>Pass/Fail</u>
Na Slope	8.70 - 11.80	Pass
BioPro Ctl L 1	70.6 - 86.2	Pass
BioPro Ctl L 3	164.7 - 201.3	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

* Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

3/20/2026
Date