



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 100 Sample
Lot Number 26049067

Ref Number 53104
Exp Date 2027-07-21
Date of Manufacture 2026-01-21
UDI# (01)00385480531047(11)260121(17)270721(10)26049067

Cal A	L/N	26036046		
		<u>Assigned Value*</u>	<u>Result</u>	
	pH	7.356	N/A	
	pCO2 mmHg	25.1	N/A	

Cal B	L/N	26036047		
		<u>Spec</u>	<u>Pass/Fail</u>	
	pH	6.837 - 6.843	Pass	
	pO2 mmHg	167.0	N/A	

Flush	L/N	26021032		
		<u>Assigned Value*</u>	<u>Result</u>	
	pCO2 mmHg	52.4	N/A	
	pO2 mmHg	103.8	N/A	

Reference	L/N	26023013		
		<u>Spec</u>	<u>Pass/Fail</u>	
	KCl	1.8 - 2.2	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector
Title

03/12/2026
Date