CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 300 Sample
Lot Number 26062018

Ref Number 53359
Exp Date 2027-08-18
Date of Manufacture 2026-02-18
UDI# (01)00385480533591(11)260218(17)270818(10)26062018

Cal A L/N 26056008

	Assigned Value*	Result
pH	7.357	N/A
pCO2 mmHg	25.0	N/A

Cal B L/N 26051033

	Spec	Pass/Fail
pH	6.837 - 6.843	Pass
pO2 mmHg	167.0	N/A

Flush L/N 26049033

	Assigned Value*	Result
pCO2 mmHg	52.8	N/A
pO2 mmHg	99.0	N/A

Reference L/N 26051034

	Spec	Pass/Fail
KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi.M

Approval

QC Inspector
Title

03/25/2026
Date