



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 100 Sample
Lot Number 26062018

Ref Number 53104
Exp Date 2027-08-18
Date of Manufacture 2026-02-18
UDI# (01)00385480531047(11)260218(17)270818(10)26062018

Cal A	L/N	26056008	
		<u>Assigned Value*</u>	<u>Result</u>
pH		7.357	N/A
pCO2 mmHg		25.0	N/A

Cal B	L/N	26051033	
		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.837 - 6.843	Pass
pO2 mmHg		167.0	N/A

Flush	L/N	26049033	
		<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg		52.8	N/A
pO2 mmHg		99.0	N/A

Reference	L/N	26051034	
		<u>Spec</u>	<u>Pass/Fail</u>
KCl		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi-M

Approval

QC Inspector
Title

03/25/2026
Date