



**CERTIFICATE OF ANALYSIS**

STP Prime ABG Cal Cartridge 200 Sample  
Lot Number 26062018

Ref Number 53360  
Exp Date 2027-08-18  
Date of Manufacture 2026-02-18  
UDI# (01)00385480533607(11)260218(17)270818(10)26062018

Cal A	L/N	26056008		
		<u>Assigned Value*</u>	<u>Result</u>	
	pH	7.357	N/A	
	pCO2 mmHg	25.0	N/A	

Cal B	L/N	26051033		
		<u>Spec</u>	<u>Pass/Fail</u>	
	pH	6.837 - 6.843	Pass	
	pO2 mmHg	167.0	N/A	

Flush	L/N	26049033		
		<u>Assigned Value*</u>	<u>Result</u>	
	pCO2 mmHg	52.8	N/A	
	pO2 mmHg	99.0	N/A	

Reference	L/N	26051034		
		<u>Spec</u>	<u>Pass/Fail</u>	
	KCl	1.8 - 2.2	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Auishi-M*

Approval

QC Inspector

Title

03/25/2026

Date