



CERTIFICATE OF ANALYSIS

Nipro Glucose and β -Ketone Control Solution Level 1
Lot Number 0426057301

Ref Number 11-115
Exp Date 2028.08.26

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57 - 65	63
Ketone mmol/L	0.1 - 0.2	0.1

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in cursive script, appearing to read "Sandang".

Approval

QC Inspector

Title

3/24/2026

Date