



**CERTIFICATE OF ANALYSIS**

STP pHox Ultra / CCX pH Sensor  
Lot Number 26082076

Part Number 48580  
Exp Date 2027-09-23  
Date of Manufacture 2026-03-23  
UDI# (01)00385480485807(11)260323(17)270923(10)26082076

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.20 - 11.20	Pass
Std C mV	0.00 ± 110.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Lumana Achta".

Approval

QC Inspector  
Title

4/9/2026  
Date