CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 200 Sample
Lot Number 26082068

Ref Number 53360
Exp Date 2027-09-02
Date of Manufacture 2026-03-02
UDI# (01)00385480533607(11)260302(17)270902(10)26082068

Cal A L/N 26061016

	Assigned Value*	Result
pH	7.355	N/A
pCO2 mmHg	24.9	N/A

Cal B L/N 26061017

	Spec	Pass/Fail
pH	6.837 - 6.843	Pass
pO2 mmHg	167.0	N/A

Flush L/N 26061013

	Assigned Value*	Result
pCO2 mmHg	52.8	N/A
pO2 mmHg	99.1	N/A

Reference L/N 26061015

	Spec	Pass/Fail
KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Aurishi.M

Approval

QC Inspector
Title

04/07/2026
Date