



## CERTIFICATE OF ANALYSIS

Stat Profile pHox Ultra / CCX BUN Membrane Kit  
Lot Number 26086015

Part Number 48596  
Exp Date 2027-06-27

Date of manufacture 2026-03-27

UDI# (01)00385480485968(11)260327(17)270627(10)26086015

	<u>Spec</u>	<u>Mean Result</u>
BUN Slope	9.30 - 12.90	12.44
CCX L 4	14 - 24	19
CCX L 5	49 - 63	52

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Lorenna A. S. Silva".

Approval

QC Inspector  
Title

4/7/2026  
Date