



## CERTIFICATE OF ANALYSIS

STP pHox Ultra / CCX pH Sensor  
Lot Number 26084032

Part Number 48580  
Exp Date 2027-09-25  
Date of Manufacture 2026-03-25  
UDI# (01)00385480485807(11)260325(17)270925(10)26084032

|          | <u>Spec</u>   | <u>Pass/Fail</u> |
|----------|---------------|------------------|
| Slope    | 9.20 - 11.20  | Pass             |
| Std C mV | 0.00 ± 110.00 | Pass             |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to read "Sumana Achta".

Approval

QC Inspector  
Title

4/9/2026  
Date