



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 CDV Reagent Pack
Lot Number 26096051

Ref Number 55297
Exp Date 2027-02-27
Date of Manufacture 2026-02-27

	CDV Bottle Pack L/N:	26096049	Exp Date:	2027-02-27		
Soln SF L/N	26064021		Soln DP L/N	26064023		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>	
pH	6.900 - 7.100	Pass	Depro Soln	N/A	Present	
Cal C L/N	26064022		Soln T L/N	26058039		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>	
3% Detergent	N/A	Present	0.4% Trypan	N/A	Present	

CDV Reagent Cartridge L/N: 26096048 Exp Date: 2027-06-11
 Soln PF L/N 26070012
 Spec Pass/Fail
 pH 6.900 - 7.100 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

4/29/2026
Date