



**CERTIFICATE OF ANALYSIS**

STAT PROFILE pHox Ultra/CCX Potassium Sensor  
Lot Number 26079059

Ref Number 48582

Exp Date 2027-06-20

Date of Manufacture 2026-03-20

UDI# (01)00385480485821(11)260320(17)270620(10)26079059

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.00 to 11.00	Pass
Std C mV	0 ± 50.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

4/7/2026  
Date