



**CERTIFICATE OF ANALYSIS**

STP pHox Ultra / CCX Lactate Sensor  
Lot Number 26083014

Part Number 48553  
Exp Date 2028-03-24  
Date of Manufacture 2026-03-24  
UDI# (01)00385480485531(11)260324(17)280324(10)26083014

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	11.00 - 45.00	Pass
pHOx Ctl L3	77 - 91	Pass
Set A L1	≥ 470	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

4/6/2026  
Date