



CERTIFICATE OF ANALYSIS

BioProfile CDV Reagent Pack
Lot Number 26082075

Ref Number 48796
Exp Date 2027-02

	CDV Bottle Pack L/N	26082074	Exp Date	2027-02		
	C Soln L/N	26076034		T Soln L/N	26071032	
		<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>
3% Detergent		N/A	Present	0.4% Trypan	N/A	Present
pH		> 9.0	Pass	Particle Count	≤ 2.0	Pass
	CDV Cartridge L/N	26082073	Exp Date	2027-03		
	S Soln L/N	26078028		D Soln L/N	26077007	
		<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Result</u>
pH		6.900 - 7.100	Pass	%NaOCl	0.43 - 0.73	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi-M

Approval

QC Inspector

Title

4/14/2026

Date