



CERTIFICATE OF ANALYSIS

STAT PROFILE pHox Ultra/CCX pCO2 Sensor
Lot Number 26097094

Ref Number 48600
Exp Date 2028-04-07
Date of Manufacture 2026-04-07
UDI# (01)00385480486002(11)260407(17)280407(10)26097094

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.40 - 11.60	Pass
Std C mV	0.00 ± 150.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "J. Smith", written over a horizontal line.

Approval

QC Inspector
Title

4/23/2026
Date