



CERTIFICATE OF ANALYSIS

Nova StatStrip Hb/Hct Control Level 2
Lot Number 2026058242

Ref Number 59827
Exp Date 2028-02-27
Date of Manufacture 2026-02-27
UDI# (01)00385480598279(11)260227(17)280227(10)2026058242

	<u>Spec</u>	<u>Mean Result</u>
Hb g/dL	14.2 - 15.8	15.3
Hct %	42 - 47	44

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awish-M

Approval

QC Inspector
Title

4/1/2026
Date