



## CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 pH/Gas Reference Sensor  
Lot Number 26085028

Part Number 57467  
Exp Date 2027-09-26  
Date of Manufacture 2026-03-26

	<u>Spec*</u>	<u>Pass/Fail</u>
pH Slope	8.70 - 11.80	Pass
BioPro Ctl L 1	6.628 - 6.728	Pass
BioPro Ctl L 3	7.328 - 7.428	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

\* Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in black ink, appearing to be "John", written over a horizontal line.

Approval

QC Inspector  
Title

04/24/2026  
Date