



CERTIFICATE OF ANALYSIS

Allegro Lipids Test Cartridge
Lot Number 26084043

Ref Number 54652
Exp Date 2027-09-25
Date of Manufacture 2026-03-25
UDI# (01)00385480546522(11)260325(17)270925(10)26084043

	Ctl Level 1				Ctl Level 2				
	<u>Spec</u>			<u>Pass/Fail</u>	<u>Spec</u>			<u>Pass/Fail</u>	
HDL mg/dL	32	-	54	Pass	HDL mg/dL	70	-	104	Pass
TCHOL mg/dL	145	-	215	Pass	TCHOL mg/dL	267	-	383	Pass
TGL mg/dL	56	-	132	Pass	TGL mg/dL	314	-	454	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector
Title

4/15/2026
Date