



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Lactate Membrane
Lot Number 26114004

Part Number 48554
Exp Date 2027-07-24
Date of Manufacture 2026-04-24
UDI# (01)00385480485548(11)260424(17)270724(10)26114004

	<u>Spec*</u>	<u>Mean Result</u>
Slope	≥ 15.00	33.91
pHox Ultra/CCX Ctl L4	1.7 - 2.3	2.0
pHox Ultra/CCX Ctl L5	6.3 - 7.7	7.1
Lin Soln Set A Level 1	14.0 - 20.0	18.2

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

5/1/2026
Date