



CERTIFICATE OF ANALYSIS

NCRT 4 Calibrator Pack
Lot Number 610750

Ref Number 15647
Exp Date 2027-07-02
Date of Manufacture 2026-04-02
UDI#(01)00385480156479(11)260402(17)270702(10)61075

CAI A L/N 26103009			Cal B L/N 26103010		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	129.1-130.9	Pass	Na mmol/L	59.4-60.6	Pass
K mmol/L	3.97-4.03	Pass	K mmol/L	9.8-10.2	Pass
Cl mmol/L	115.1-116.9	Pass	Cl mmol/L	60.4-61.6	Pass
HCO3 mmol/L	19.2-20.8	Pass	HCO3 mmol/L	9.6-10.4	Pass
Li mmol/L	0.50-0.51	Pass	Li mmol/L	4.95-5.05	Pass

Cal H L/N 26097017			Reference L/N 26092014		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Pass/Fail</u>
CO2 Releasing Agent	N/A	Present	KCl mol/L	1.8-2.2	Pass

accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

4/24/2026
Date