



CERTIFICATE OF ANALYSIS

Nova Cleaning Solution
Lot Number 26112012

Ref Number 11272
Exp Date 2027-04-22
Date of Manufacture 2026-04-22
UDI# (01)00385480112727(11)260422(17)270422(10)26112012

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	1.21 - 1.51%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in blue ink, appearing to read "C. A. V. E. T.", written over a horizontal line.

Approval

QC Inspector
Title

5/7/2026
Date