



CERTIFICATE OF ANALYSIS

Nova 16 Calibrator Pack
Lot Number 611326

Ref Number 17289
Exp Date 2027-01-23
Date of Manufacture 2026-04-23
UDI# (01)00385480172899(11)260423(17)270123(10)611326

Cal A L/N 26113026			Cal B L/N 26113026			Cal C L/N 26113026		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>
HCO3 mmol/L	19.2 - 20.8	Pass	HCO3 mmol/L	19.6 - 20.8	Pass	HCO3 mmol/L	9.6 - 10.4	Pass
Na mmol/L	139.0 - 141.0	Pass	Na mmol/L	79.2 - 80.8	Pass			
K mmol/L	3.97 - 4.03	Pass	K mmol/L	9.8 - 10.2	Pass			
Cl mmol/L	112.1 - 113.9	Pass	Cl mmol/L	57.4 - 58.6	Pass			
Glu mg/dL	78.4 - 81.6	Pass	Glu mg/dL	196 - 204	Pass			
BUN mg/dL	9.7 - 10.3	Pass	BUN mg/dL	49.4 - 50.6	Pass			
Creat mg/dL	0.98 - 1.02	Pass	Creat mg/dL	9.8 - 10.2	Pass			

Flush L/N 26113026			Reference L/N 26113026		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>
Aqueous Salt Solution	N/A	Present	KCl mol/L	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

5/13/2026
Date