



CERTIFICATE OF ANALYSIS

STAT PROFILE® pHOx Ultra™/CCX™ pCO2 Sensor
Lot Number 22193009

Ref Number 48600
Exp Date 2024-07-12
Date of Manufacture 2022-07-12
UDI# (01)00385480486002(11)220712(17)240712(10)22193009

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.40 - 11.60	Pass
Std C mV	0.00 ± 150.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

8/11/2022
Date