



CERTIFICATE OF ANALYSIS

Nova 1/5/8/11/13/14 Hematocrit Sensor
Lot Number 26089025

Part Number 13049
Exp Date 2027-03-30
Date of Manufacture 2026-03-30
UDI# (01)00385480130493(11)260330(17)270330(10)26089025

	<u>Spec*</u>	<u>Pass/Fail</u>
Hct Slope	8.0 - 20.0	Pass
Set D1	13 - 17	Pass
Set D2	62 - 68	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi.M

Approval

QC Inspector

Title

6/1/2026

Date