



**CERTIFICATE OF ANALYSIS**

Linearity Std Set A Level 6  
Lot Number 26148013

Ref Number 14071  
Exp Date 2027-11-28  
Date of Manufacture 2026-05-28  
UDI# (01)00385480140713(11)260528(17)271128(10)26148013

	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	98.0 - 102.0	Pass
K+ mmol/L	98.0 - 102.0	Pass
Cl- mmol/L	196.0 - 204.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

*Awishi.M*

Approval

QC Inspector  
Title

6/12/2026  
Date