



## CERTIFICATE OF ANALYSIS

NOVA PRO Glucose and Ketone Control Solution Level 2  
Lot Number 7526118302

Ref Number 52332  
Exp Date 2028-10-28  
Date of Manufacture 2026-04-28  
UDI# (01)00385480523325(11)260428(17)281028(10)7526118302

|               | <u>Spec</u> | <u>Mean Result</u> |
|---------------|-------------|--------------------|
| Glucose mg/dL | 98 - 128    | 119                |
| Ketone mmol/L | 0.6 - 0.8   | 0.7                |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Awishi-M*

Approval

QC Inspector  
Title

6/15/2026  
Date