



CERTIFICATE OF ANALYSIS

STP pHox pH Conditioning Solution
Lot Number 26163027

Ref Number 23397
Exp Date 2027-09-12

Date of Manufacture 2026-06-12

UDI# (01)00385480233972(11)260612(17)270912(10)26163027

	<u>Spec</u>	<u>Pass/Fail</u>
pH	≤ 5.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. M. ...", written over a light blue horizontal line.

Approval

QC Inspector
Title

6/18/2026
Date