



CERTIFICATE OF ANALYSIS

BioProfile Control Level 8
Lot Number 26159016

Ref Number 43472
Exp Date 2027-09

	<u>Spec</u>	<u>Pass/Fail</u>
Particles/mL	$5.50 \times 10^5 - 6.50 \times 10^5$	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi.M

Approval

QC Inspector
Title

6/22/2026
Date