



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Lactate Membrane
Lot Number 26161008

Part Number 48554
Exp Date 2027-09-10
Date of Manufacture 2026-06-10
UDI# (01)00385480485548(11)260610(17)270910(10)26161008

	<u>Spec*</u>	<u>Mean Result</u>
Slope	≥ 15.00	30.36
pHOx Ultra/CCX Ctl L4	1.7 - 2.3	2.0
pHOx Ultra/CCX Ctl L5	6.3 - 7.7	7.4
Lin Soln Set A Level 1	14.0 - 20.0	19.0

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Lumina Fichta".

Approval

QC Inspector
Title

6/18/2026
Date