



CERTIFICATE OF ANALYSIS

Nova Glu/Lac Sensor Soaking Solution
Lot Number 26162004

Ref Number 13409

Exp Date 2027-06-11

Date of Manufacture 2026-06-11

UDI# (01)00385480134095(11)260611(17)270611(10)26162004

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	0.43 - 0.73%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Auishi.M

Approval

QC Inspector
Title

6/26/2026
Date