



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 500 Sample
Lot Number 26155057

Ref Number 53464
Exp Date 2027-11-20
Date of Manufacture 2026-05-20
UDI# (01)00385480534642(11)260520(17)271120(10)26155057

Cal A	L/N	26140005		
			<u>Assigned Value*</u>	<u>Result</u>
pH			7.361	N/A
pCO2 mmHg			24.3	N/A

Cal B	L/N	26140006		
			<u>Spec</u>	<u>Pass/Fail</u>
pH			6.837 - 6.843	Pass
pO2 mmHg			167.0	N/A

Flush	L/N	26140002		
			<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg			52.8	N/A
pO2 mmHg			96.0	N/A

Reference	L/N	26140041		
			<u>Spec</u>	<u>Pass/Fail</u>
KCl			1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi-M

Approval

QC Inspector

Title

06/22/2026

Date