



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 CDV Reagent Pack
Lot Number 26166088

Ref Number 55297
Exp Date 2027-05-13
Date of Manufacture 2026-05-13

CDV Bottle Pack L/N:	26166087	Exp Date:	2027-05-13
Soln SF L/N	26135024	Soln DP L/N	26134032
	<u>Spec</u>		<u>Spec</u>
	<u>Result</u>		<u>Result</u>
pH	6.900 - 7.100	Depro Soln	N/A
	Pass		Present
Cal C L/N	26134046	Soln T L/N	26133030
	<u>Spec</u>		<u>Spec</u>
	<u>Result</u>		<u>Result</u>
3% Detergent	N/A	0.4% Trypan	N/A
	Present		Present

CDV Reagent Cartridge L/N: 26166085 Exp Date: 2027-08-27
Soln PF L/N 26147034
 Spec Pass/Fail
 pH 6.900 - 7.100 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

6/29/2026
Date