



CERTIFICATE OF ANALYSIS

Nova 1/4/5/8/10/11/12/13/14 Reference Sensor
Lot Number 26177034

Part Number 14244
Exp Date 2027-12-26
Date of Manufacture 2026-06-26
UDI# (01)00385480142441(11)260626(17)271226(10)26177034

	<u>Spec*</u>	<u>Pass/Fail</u>
Na Slope	≥ 9.8	Pass
Set A Level 3	137.2 - 143.2	Pass
CI Slope	≥ 10.5	Pass
Set A Level 3	110.4 - 118.4	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

²Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

7/2/2026

Date